

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Benzene



## SBE-2M

TWA (TLV): 0.5 ppm

STEL (TLV): 2.5 ppm

Flammable Range: 1-8%

### Specification:

	Extended	Standard	Extended
Measurement Range	1 – 75 ppm	0.2 - 15 ppm	0.1-7.5 ppm
No. of Pump Strokes	1 (100 mL)	5 (500 mL)	10 (1000 mL)
Volume Correction Factor (VCF)*	5	1.0	0.5
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White → Brown		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_6H_6 + I_2O_5 + H_2S_2O_7 \rightarrow I_2$		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Toluene	5	Positive	Brown/Higher reading
Xylene	15	Positive	Brown/Higher reading
CO	10	Positive	Brown/Higher reading
Hexane	15	Positive	Brown/Higher reading

### Correction for Environmental Parameters:

Temperature (°C) :	10	20	30	40
Temperature Correction Factor (TCF) *:	1.09	1	0.95	0.89

\* Multiply the observed reading by the correction factor (VCFXTCF) to obtain the true concentration.

Relative humidity	Not necessary between 10 - 90 %.
-------------------	----------------------------------



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.1115