

# Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

## Butyl Acetate

$C_6H_{12}O_2$

## SBA-6

TWA (TLV): 150 ppm

STEL (TLV): 200 ppm

Flammable Range: 1.7 – 7.6 %

### Specification:

Measurement Range	Extended	Standard	Extended
	Do not extend	0.005 – 1 %	Do not extend
No. of Pump Strokes		1 (100 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	Yellow → Brownish green → Green		
Shelf Life	2 years		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$CH_3CO_2(CH_2)_3CH_3 + Cr^{6+} + H_2SO_4 \rightarrow Cr^{3+}$		
Relative standard deviation	± 10 %		

\* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

### Possible Interferences:

Compound	Concentration (%)	Interference	Colour change / Comments
Alcohol / esters/ ketones		+	
Propane	0.05	+	Yellow - Brownish green
H <sub>2</sub> S	0.1	+	Diffuse boundary
SO <sub>2</sub>	0.1	+	Ring

### Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



**Manufactured By:** Uniphos Envirotronic Pvt. Ltd.  
Nahuli (PO), Umbergaon (Tal.)  
Valsad (Dist), Gujarat, India – 396108

Rev.0812