

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

1,2-Dichlorobenzene $C_6H_4Cl_2$ SDB-4M

TWA (TLV): 0.1 mg/L

STEL (TLV): NA

Flammable Range: 0.5-6%

Specification:

	Extended	Standard	Extended
Measurement Range	Do not extend	10 – 300 ppm	Do not extend
No. of Pump Strokes		2 (200 mL)	
Volume Correction Factor (VCF)*		1.0	
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	White → Brown		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F) and 50% RH.		
Reaction Principle	$C_6H_4Cl_2 + I_2O_5 + H_2SO_7 \rightarrow I_2$		
Relative standard deviation	± 10 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
Benzene		+	White to Brown
Toluene		+	White to Brown
CO		+	White to Brown
Hexane	1100	No	No Effect

Correction for Environmental Parameters:

Temperature	Not necessary between 0 - 40°C (32 - 104°F).
Relative humidity	Not necessary between 10 - 90 %.



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0812