

Uniphos Detector Tube Data Sheet

Direct Reading Length of Stain Type

Water Vapor

H₂O

SWA-4L

TWA (TLV): NA

STEL (TLV): NA

Non Flammable

Specification:

	Extended	Standard	Extended
Measurement Range	1 – 5 lb/MMCF	2 – 10 lb/MMCF	4 – 20 lb/MMCF
No. of Pump Strokes	4 (400 mL)	2 (200 mL)	01 (100 mL)
Volume Correction Factor (VCF)*	0.5	1.0	2.0
Sampling Time	1.5 minutes per pump stroke (100 mL)		
Color Change	Yellow → Pale green		
Shelf Life	1 year		
Storage condition	Below 10°C (50°F)		
Calibration condition	Calibrated at 20°C (68°F).		
Reaction Principle	H ₂ O + Mg (ClO ₄) ₂ → Mg(ClO ₄) ₂ .H ₂ O		
Relative standard deviation	± 10 %		

* Multiply the observed reading by the correction factor (VCF) to obtain the true concentration.

Possible Interferences:

Compound	Concentration (ppm)	Interference	Colour change / Comments
CH ₄	99.9%	No effect	No effect
H ₂ S	2000	No effect	No effect
Methanol	50	No effect	No effect

Correction for Environmental Parameters:

Temperature	Temperature correction is necessary.
-------------	--------------------------------------



Manufactured By: Uniphos Envirotronic Pvt. Ltd.
Nahuli (PO), Umbergaon (Tal.)
Valsad (Dist), Gujarat, India – 396108

Rev.0812